

# United States Patent and Trademark Office

UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER OF PATENTS AND TRADEMARKS Washington, D.C. 20231 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.	
09/776,125	02/02/2001	Seth A. Foerster	END-723	1306	
75	90 07/03/2002				
Audley A. Ciamporcero, Jr., Esq. Johnson & Johnson One Johnson & Johnson Plaza			EXAMINER		
			BAXTER, JESSICA R		
New Brunswick, NJ 08933-7003			ART UNIT	PAPER NUMBER	
			3731	· · · · · · · · · · · · · · · · · · ·	
			DATE MAILED: 07/03/2002	DATE MAILED: 07/03/2002	

Please find below and/or attached an Office communication concerning this application or proceeding.

ŧ		•	55			
Office Action Summary		Application No.	Applicant(s)			
		09/776,125	FOERSTER ET AL.			
		Examiner	Art Unit			
		Jessica R Baxter	3731			
Period fo	- The MAILING DATE of this communication a r Reply	ppears on the cover shee	t with the correspondence address			
THE N - Extens after S - If the I - If NO - Failure - Any re	PRTENED STATUTORY PERIOD FOR REF MAILING DATE OF THIS COMMUNICATION sions of time may be available under the provisions of 37 CFR sions of time may be available under the provisions of 37 CFR (6) MONTHS from the mailing date of this communication. Deriod for reply specified above is less than thirty (30) days, a reperiod for reply sis specified above, the maximum statutory perion to reply within the set or extended period for reply will, by state the province of the provinc	N. 1.136(a). In no event, however, ma eply within the statutory minimum o od will apply and will expire SIX (6) tute, cause the application to becom	y a reply be timely filed  thirty (30) days will be considered timely.  MONTHS from the mailing date of this communication. e ABANDONED (35 U.S.C. § 133).			
1)⊠	Responsive to communication(s) filed on $\underline{o}$	8 August 2001 .				
2a) <u></u> □	This action is <b>FINAL</b> . 2b)⊠	This action is non-final.				
3) 🗌	Since this application is in condition for allo					
Disposition	closed in accordance with the practice undo on of Claims	ei Ex parte Quayle, 1955	C.D. 11, 433 C.G. 213.			
4) 🖂	Claim(s) <u>1-10</u> is/are pending in the applicati	ion.				
4	a) Of the above claim(s) is/are withd	rawn from consideration.				
5)	Claim(s) is/are allowed.					
6)⊠	Claim(s) <u>1-10</u> is/are rejected.					
7)	Claim(s) is/are objected to.					
• —	Claim(s) are subject to restriction and	d/or election requirement.				
	on Papers					
•	he specification is objected to by the Exami					
10)∐ T	10) ☐ The drawing(s) filed on is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.					
44) 🗀 🕇	Applicant may not request that any objection to	- · ·				
11)[_] 1	he proposed drawing correction filed on If approved, corrected drawings are required in		disapproved by the Examiner.			
12\□ T	The oath or declaration is objected to by the					
•	nder 35 U.S.C. §§ 119 and 120	Examinor.				
-	Acknowledgment is made of a claim for fore	ian priority under 35 H S	C & 119(a)-(d) or (f)			
•	Acknowledgment is made of a dailin for force  All b) Some * c) None of:	igh phoney under 55 C.C.	o. 9 113(a)-(u) or (i).			
• —	1.☐ Certified copies of the priority docume	ents have been received				
	2. ☐ Certified copies of the priority docume		n Application No			
	3. ☐ Copies of the certified copies of the pr					
	application from the International I ee the attached detailed Office action for a li	Bureau (PCT Rule 17.2(a	)).			
14)∐ A	cknowledgment is made of a claim for dome	stic priority under 35 U.S	C. § 119(e) (to a provisional application).			
•	☐ The translation of the foreign language pcknowledgment is made of a claim for dome	• •				

U.S. Patent and Trademark Office PTO-326 (Rev. 04-01)

1) Notice of References Cited (PTO-892)

2) Notice of Draftsperson's Patent Drawing Review (PTO-948)

3) Information Disclosure Statement(s) (PTO-1449) Paper No(s) 5.

Attachment(s)

4) Interview Summary (PTO-413) Paper No(s). \_

5) Notice of Informal Patent Application (PTO-152)

#### **DETAILED ACTION**

## Specification

1. The term "various means" is used. Applicant is reminded of the proper language and format for an abstract of the disclosure.

The abstract should be in narrative form and generally limited to a single paragraph on a separate sheet within the range of 50 to 150 words. It is important that the abstract not exceed 150 words in length since the space provided for the abstract on the computer tape used by the printer is limited. The form and legal phraseology often used in patent claims, such as "means" and "said," should be avoided. The abstract should describe the disclosure sufficiently to assist readers in deciding whether there is a need for consulting the full patent text for details.

The language should be clear and concise and should not repeat information given in the title. It should avoid using phrases which can be implied, such as, "The disclosure concerns," "The disclosure defined by this invention," "The disclosure describes," etc.

### Claim Objections

2. The numbering of claims is not in accordance with 37 CFR 1.126 which requires the original numbering of the claims to be preserved throughout the prosecution. When claims are canceled, the remaining claims must not be renumbered. When new claims are presented, they must be numbered consecutively beginning with the number next following the highest numbered claims previously presented (whether entered or not).

Misnumbered claims 15-17, 35-37 and 43-45 have been renumbered 2-10. Claim Rejections - 35

### USC § 112

3. The following is a quotation of the second paragraph of 35 U.S.C. 112:

The specification shall conclude with one or more claims particularly pointing out and distinctly claiming the subject matter which the applicant regards as his invention.

Application/Control Number: 09/776,125

Art Unit: 3731

4. Claim 43 is rejected under 35 U.S.C. 112, second paragraph, as being indefinite for failing to particularly point out and distinctly claim the subject matter which applicant regards as the invention. Claim 43 improperly depends upon claim 38 which is not pending in this application.

### Claim Rejections - 35 USC § 103

- 1. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:
  - (a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.
- 2. Claims 1-3 and 5-9 are rejected under 35 U.S.C. 103(a) as being unpatentable over U.S. Patent No. 3,958,576 to Komiya in view of U.S. Patent no. 4,929,240 to Kirsch et al. and U.S. Patent No. 5,221,269 to Miller et al. Komiya shows a device comprising: a discrete marker element (11), a member (10), a deployment actuator connector (15/16/21), a forming die (21b) wherein the marker element is able to travel along the tube, but Komiya does not show the cutting tip and the predetermined failure point. Komiya further describes the use of known delivery apparatuses, such as endoscopes, to place the device into proper position (see Column 2 lines 47-49). Applicant's summary discloses that the delivery apparatus may be any known type such as a tube, a needle, a cannula or a trocar. Miller shows an implantable tissue-marking device using a delivery apparatus that incorporates a needle or cutting edge (see Column 4 lines 63-68). It would have been obvious to one having ordinary skill in the art at the time the invention was made to have used a needle as taught in Miller because such a variation is an obvious art-recognized alternative, as even disclosed by the applicant, to serve the purpose of placing the marking device in the proper location in the

body. Kirsch shows a deployment actuator having a predetermined failure point. It would have been obvious to one having ordinary skill in the art at the time the invention was made to use the deployment means of Kirsch because both structures are considered art-recognized alternative structures. Both incorporate bendable marker elements that are similar in shape, but use different styles of deployment means, Komiya uses a notch and Kirsch breaks the marker off. Regarding the limitations drawn to various marker shapes or sizes, they are considered to be obvious choices of design based on user preference to suit a particular need or application.

## Double Patenting

3. The nonstatutory double patenting rejection is based on a judicially created doctrine grounded in public policy (a policy reflected in the statute) so as to prevent the unjustified or improper timewise extension of the "right to exclude" granted by a patent and to prevent possible harassment by multiple assignees. See *In re Goodman*, 11 F.3d 1046, 29 USPQ2d 2010 (Fed. Cir. 1993); *In re Longi*, 759 F.2d 887, 225 USPQ 645 (Fed. Cir. 1985); *In re Van Ornum*, 686 F.2d 937, 214 USPQ 761 (CCPA 1982); *In re Vogel*, 422 F.2d 438, 164 USPQ 619 (CCPA 1970);and, *In re Thorington*, 418 F.2d 528, 163 USPQ 644 (CCPA 1969).

A timely filed terminal disclaimer in compliance with 37 CFR 1.321(c) may be used to overcome an actual or provisional rejection based on a nonstatutory double patenting ground provided the conflicting application or patent is shown to be commonly owned with this application. See 37 CFR 1.130(b).

Effective January 1, 1994, a registered attorney or agent of record may sign a terminal disclaimer. A terminal disclaimer signed by the assignee must fully comply with 37 CFR 3.73(b).

- 4: Claims 1, 16, 44 and 45 are rejected under the judicially created doctrine of obviousness-type double patenting as being unpatentable over claims 1 and 2 of U.S. Patent No. 6,228,055. Although the conflicting claims are not identical, they are not patentably distinct from each other because both claim a device that includes a discrete marker element, a delivery apparatus, a member that receives a deployment actuator connector, a deployment actuator, a predetermined failure point, a cutting tip, and both may be used in combination with a medical instrument that employs a vacuum.
- 5. A rejection based on double patenting of the "same invention" type finds its support in the language of 35 U.S.C. 101 which states that "whoever invents or discovers any new and useful process ... may obtain a patent therefor ..." (Emphasis added). Thus, the term "same invention," in

Application/Control Number: 09/776,125

Art Unit: 3731

this context, means an invention drawn to identical subject matter. See *Miller v. Eagle Mfg. Co.*, 151 U.S. 186 (1894); *In re Ockert*, 245 F.2d 467, 114 USPQ 330 (CCPA 1957); and *In re Vogel*, 422 F.2d 438, 164 USPQ 619 (CCPA 1970).

A statutory type (35 U.S.C. 101) double patenting rejection can be overcome by canceling or amending the conflicting claims so they are no longer coextensive in scope. The filing of a terminal disclaimer <u>cannot</u> overcome a double patenting rejection based upon 35 U.S.C. 101.

6. Claim 17 is rejected under 35 U.S.C. 101 as claiming the same invention as that of claim 1 of prior U.S. Patent No. 6,228,055. This is a double patenting rejection.

#### Conclusion

7. This is a continuation of applicant's earlier Application No. 08/858389. All claims are drawn to the same invention claimed in the earlier application and could have been finally rejected on the grounds and art of record in the next Office action if they had been entered in the earlier application. Accordingly, **THIS ACTION IS MADE FINAL** even though it is a first action in this case. See MPEP § 706.07(b). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no, however, event will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Jessica R Baxter whose telephone number is 703-305-4069. The examiner can normally be reached on M-F 8:30AM - 5:00PM.

Application/Control Number: 09/776,125

Art Unit: 3731

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Michael Milano can be reached on 703-308-2496. The fax phone numbers for the organization where this application or proceeding is assigned are 703-305-3590 for regular communications and 703-305-3590 for After Final communications.

Any inquiry of a general nature or relating to the status of this application or proceeding should be directed to the receptionist whose telephone number is 703-308-0858.

essica R Baxter Examiner

Art Unit 3731

jrb June 28, 2002

> MICHAEL J. MILANO SUPERVISORY PATENT EXAMINER TECHNOLOGY CENTER 3700